



A Company of Sysmex Group



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PT-Phen™ LRT
Product Code : CK586K-CK586L
Lot No. : FD1591M7
Expiration date : 2026-09-12

Components	Quantity	Lot	Expiration Date
R1 : PT-Phen™ LRT	6 vials (CK586K) 12 vials (CK586L)	FD1591doM7	2026-09-12

Analytical Data	Specification	Result
PT on Normal Plasma	11 ± 1 sec	12.0 sec (On CS)
ISI (WHO Method)	1.0 ± 0.1	0.94 on CS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

17 JUL. 2025

S. LECOURT