



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN EDOXABAN CALIBRATOR LOW
Product Code : 226401
Lot No. : FD10101B
Expiration date : 2026-11-10

Components	Quantity	Lot	Expiration Date
CAL I	4 vials	FD1010af1B	2026-11-10
CAL II	4 vials	FD1010ag1B	2026-11-10
CAL III	4 vials	FD1010ah1B	2026-11-10

Analytical Data	Specification	Result
Concentration	CAL I : < 15 ng/ml CAL II : 40 -60 ng/ml CAL III : 80 -120 ng/ml	0 ng/ml 50 ng/ml 103 ng/ml
Linearity and accuracy of calibration with values assigned	$R^2 \geq 0.98$ Measured values for controls in acceptance ranges	PASS PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** Quality Control Manager :

11 JUN 2024
S. LECOURT

