

 HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ EDOXABAN CONTROL
Product Code : 225501
Lot No. : FD18421E
Expiration date : 2027-03-15

Components	Quantity	Lot	Expiration Date
C1	6 vials	FD1842ad1E	2027-03-15
C2	6 vials	FD1842ae1E	2027-03-15

Analytical Data	Specification	Result
Concentration	C1 : 100 -200 ng/ml C2 : 250 -350 ng/ml	135 ng/ml 277 ng/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Quality Control Approval
Date of Approval : Quality Control Manager :

10 OCT. 2024

S. LECOURT

