

A Company of Sysmex Group



HYPHEN BioMed 155 rue d'Eragny 95000 Neuville-sur-Oise France

CERTIFICATE OF ANALYSIS

Product Description:

BIOPHEN Calibrator Factor VIIa

Product Code:

226301

Lot No.:

FE205719 / FE212015

Expiration date :

2027-09-03

Components	Quantity	Lot	Expiration Date
Calibrator VIIa	6 vials	FE2057an19 / FE212015	2027-09-03

Analytical Data	Specification	Result
Concentration	350-500 mIU/mI	450 mIU/mI
Linearity and accuracy of calibration with values assigned	$R^2 \ge 0.98$ Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval:

Quality Control ApprovaQuality Control Manager:

0 3 AVR. 2025

S. LECOURT