


 HYPHEN BioMed
 155 rue d'Eragny
 95000 Neuville-sur-Oise
 France

CERTIFICATE OF ANALYSIS

Product Description : RISTOCETIN
Product Code : AG004K
Lot No. : FE2330PP / FE2354
Expiration date : 2027-10-05

Components	Quantity	Lot	Expiration Date
RISTOCETIN	3 vials	FE2330PP / FE2354	2027-10-05

Analytical Data	Specification	Result
Reactivity (% AgregMax) at 1.2mg/ml on CS or CN series	≥ 50%	74 %
Reactivity (% AgregMax) at 1.2mg/ml on Agregometer	≥ 50%	81%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**
 02 JUIN 2025
 S. LECOURT

Quality Control Manager :

