



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN LMWH CONTROL PLASMA
Product Code : 223001
Lot No. : FE2631 / FE2654
Expiration date : 2027-11-14

Components	Quantity	Lot	Expiration Date
C3	6 vials	FE2630BOQG / FE2652	2027-11-14
C4	6 vials	FE2629BPQF / FE2653	2027-11-14

Analytical Data	Specification	Result
Concentration	C3 : 0.70 -0.90 IU/ml C4 : 1.10 -1.30 IU/ml	0.74 IU/ml 1.12 IU/ml

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

19 JUIN 2025

Quality Control Manager :

Claire DUNOIS
Dir. Etudes Clin. & Val.

