



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : PLASMA ARGATROBAN
CALIBRATOR
Product Code : SC030K
Lot No. : FE32481A / FE3265
Expiration date : 2028-02-08

Components	Quantity	Lot	Expiration Date
CAL 1	4 vials	FE3248BZ1A / FE3265	2028-02-08
CAL 2	4 vials	FE3248CA1A / FE3265	2028-02-08
CAL 3	4 vials	FE3248CB1A / FE3265	2028-02-08
CAL 4	4 vials	FE3248CC1A / FE3265	2028-02-08
CAL 5	4 vials	FE3248CD1A / FE3265	2028-02-08

Analytical Data	Specification	Result
CAL 1	$\leq 0.05 \mu\text{g/mL}$	0.00 $\mu\text{g/mL}$
CAL 2	0.35 - 0.65 $\mu\text{g/mL}$	0.48 $\mu\text{g/mL}$
CAL 3	0.85 - 1.15 $\mu\text{g/mL}$	1.00 $\mu\text{g/mL}$
CAL 4	1.30 - 1.65 $\mu\text{g/mL}$	1.46 $\mu\text{g/mL}$
CAL 5	1.70 - 2.00 $\mu\text{g/mL}$	1.86 $\mu\text{g/mL}$

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval**

18 SEP. 2025

S. LECOURT

Quality Control Manager :