



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : BIOPHEN™ BIVALIRUDIN CALIBRATOR
Product Code : 226701
Lot No. : FE355619 / FE3540
Expiration date : 2028-03-06

Components	Quantity	Lot	Expiration Date
CAL 1	2 vials	FE3556cz19 / FE3540	2028-03-06
CAL 2	2 vials	FE3556da19 / FE3540	2028-03-06
CAL 3	2 vials	FE3556db19 / FE3540	2028-03-06
CAL 4	2 vials	FE3556dc19 / FE3540	2028-03-06
CAL 5	2 vials	FE3556dd19 / FE3540	2028-03-06

Analytical Data	Specification	Result
Concentration	CAL 1 : < 0.25 µg/mL CAL 2 : 0.8 - 1.2 µg/mL CAL 3 : 1.6 - 2.4 µg/mL CAL 4 : 3.0 - 4.0 µg/mL CAL 5 : 4.0 - 5.0 µg/mL	0.00 µg/mL 1.04 µg/mL 2.03 µg/mL 3.36 µg/mL 4.44 µg/mL
Accuracy of calibration with values assigned	Measured values for controls in acceptance ranges	PASS

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval : **Quality Control Approval** , Quality Control Manager :

02 OCT. 2025

S. LECOURT