



HYPHEN BioMed
155 rue d'Eragny
95000 Neuville-sur-Oise
France

CERTIFICATE OF ANALYSIS

Product Description : ZYMUTEST vWF

Product Code : RK030A

Lot No. : FE2302

Expiration date : 2028-05-01

Components	Quantity	Lot	Expiration Date
Anti (h) vWF pre-coated plate	12x8 wells	FE2287	2028-09-18
vWF standard	3 vials	FE2284	2028-09-20
vWF Control I	1 vial	FE2285	2028-09-20
vWF Control II	1 vial	FE2286	2028-09-20
Anti-(h)-vWF HRP immunoconjugate	3 vials	FE2342	2028-09-27
Sample diluent	2x50 ml	FE0193	2028-07-03
Wash solution	1x50 ml	FE0011	2028-06-15
Conjugate diluent	1x25 ml	FD2352	2028-05-01
TMB	1x25 ml	240508D01	2028-05-31
Sulfuric Acid 0,45M	1x6 ml	FE0017	2028-06-19

Analytical Data	Specification	Result
Reactivity	A450 for 100 % ≥ 1.40	1.69
Calibration Curve and Accuracy	Controls Measured Values within acceptance ranges	PASS
Blank	< 0.100	0.02
Detection threshold	$\leq 5\%$	$< 1\%$

Calibration Curve for information

Conc.	C	C/2	C/4	C/10	C/20
OD	1.878	1.431	0.921	0.451	0.298

MV = Measured value

	Acceptance Range	MV
CI	[62 - 84]	75%
CII	[34 - 52]	42%

This is to certify that above product has been produced and passed the inspection according to specifications of HYPHEN BioMed.

Date of Approval :

Quality Control Approval

Quality Control Manager :

15 MAI 2025

S. LECOURT